



CALIBRATION REPORT

ORDER No.

JUNE 18, 2019

PAGE 1 OF 2

MANUFACTURER: OHM-LABS
DESCRIPTION: CURRENT SHUNT
MODEL: CSA-20
SERIAL:

PROCEDURE: CS CAL
LAB ENVIRONMENT: 23.6 °C / 46 %RH
CALIBRATION DATE: 18/JUN/2019

Table with 3 columns: APPLIED CURRENT, MEASURED VALUE, UNCERTAINTY. Rows show data for 4A, 8, 12, 16, and 20.

NOTES: SHUNT WAS ALLOWED TO FULLY STABILIZE AT EACH APPLIED CURRENT.

Table titled 'STANDARDS USED' with columns: ID, DESCRIPTION, MAKE & MODEL, CAL DUE. Lists standards AS3021, AS3403, and AS3407.

COMMENTS:

OHM-LABS, INC. CERTIFIES THAT THIS CALIBRATION IS TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY (NIST), OR ANOTHER RECOGNIZED NATIONAL MEASUREMENT INSTITUTE, OR DERIVED BY A RATIO TYPE SELF-CALIBRATION TECHNIQUE, AND IS ACCREDITED TO ISO/IEC 17025.

PERFORMED BY:
REVIEWED BY:





# CALIBRATION REPORT

ORDER NO.

JUNE 18, 2019

PAGE 2 OF 2

MANUFACTURER: OHM-LABS

MODEL: CSA-20


SERIAL:

NOTE: AC-DC DIFFERENCE WAS MEASURED AT APPROXIMATELY 100 % RATED CURRENT OF THE SHUNT.

AC-DC DIFFERENCE DATA: AS-FOUND & AS-LEFT		
APPLIED FREQUENCY	AC-DC DIFFERENCE	UNCERTAINTY
60 HZ	-10 $\mu$ A/A	70 $\mu$ A/A
400	+20	360
1 KHZ	-130	490
5	-120	100
10	-80	120
20	-100	110
50	-80	110
70	-50	100
100	-60	70

AC-DC DIFFERENCE =  $(I_{AC} - I_{DC}) / I_{DC}$ . A POSITIVE SIGN INDICATES THAT MORE AC CURRENT THAN DC CURRENT IS NECESSARY TO PRODUCE THE SAME VOLTAGE OUTPUT ON THE SHUNT UNDER TEST. THE SHUNT UNDER TEST WAS MEASURED IN SERIES WITH A STANDARD SHUNT, AND THE OUTPUTS OF THE STANDARD AND THE UUT WERE MEASURED WITH THERMAL VOLTAGE CONVERTERS.

CHART OF AC-DC DIFFERENCE IN  $\mu$ A/A VS. FREQUENCY IN KHZ



### STANDARDS USED FOR AC-DC DIFFERENCE MEASUREMENTS

ID	DESCRIPTION	MAKE & MODEL	CAL DUE
AS3821-3	STANDARD AC SHUNT	OHM-LABS CSA-20	30/MAY/2020
AS3840	STD THERMAL CONVERTER	PTB/IPHT MJTVC	16/NOV/2021
AS3841	UUT THERMAL CONVERTER	PTB/IPHT MJTVC	15/NOV/2021

END OF REPORT